

# XP Series Balances

## Operation Manual

*North and South America:*

**Denver Instrument Company**  
6542 Fig Street  
Arvada, Colorado 80004  
1-800-321-1135  
Tel: 303-431-7255  
Fax: 303-423-4831


*U.K. and Ireland:*

**Denver Instrument Company**  
Denver House, Sovereign Way  
Trafalgar Business Park  
Downham Market  
Norfolk PE38 9SW England  
Tel: 44 136 63862 42  
Fax: 44 136 63862 04

*Europe, Asia and Australia:*

**Denver Instrument GmbH**  
Robert-Bosch-Briete 10  
37079 Gottingen Germany  
Tel: 49 551 20977 31  
Fax: 49 551 20977 39

[www.denverinstrument.com](http://www.denverinstrument.com)



You have purchased a quality precision weighing instrument that requires handling with care.

*Read entire contents of this **Operation Manual** prior to operating your new Denver Instrument balance.*

## Disclaimer Notice

Calibrate your balance using a reference weight of the appropriate tolerance (class). An instrument can be no more accurate than the standard to which it has been compared. For assistance in the selection of reference weights, please contact the factory.



### Caution!

Changes or modifications not expressly approved by the manufacturer could void the user's authority to operate this equipment.



### Warning

**Never lift balance by the weighing pan as this may cause damage to internal mechanisms.**

*Always lift and transport the balance by its base, including removal from packing materials!*

### Warning

**Do not drop weights on pan.**

Damage will occur to the balance's internal mechanisms.

### Note:

Damage may occur if load exceeds 150% of maximum capacity.

### Note:

This product may contain a lead acid battery. The user is responsible for disposing of the unit in accordance with local ordinances and regulations.



6542 Fig Street • Arvada, Colorado 80004  
(303) 431-7255 • (800) 321-1135 • Fax(303) 423-4831

## Cleaning Instructions

The exterior surfaces of the product may be cleaned with a mild water-based detergent and a lint-free nonabrasive cloth. Alternatively, isopropyl alcohol may be used. Do not immerse the product in any liquid.



### Warning

**Never lift balance by the weighing pan as this may cause damage to internal mechanisms.**

*Always lift and transport the balance by its base.*

### Note

***This unit contains no user serviceable parts. All replacement parts should be obtained from the manufacturer. Please refer to this manual for the phone number of your sales and service representative.***

## Warranty Instructions

1. Please return the prepaid, pre-addressed Purchase Registration Card to Denver Instrument Company promptly upon your purchase of the Denver Instrument product. The return of the card is not a condition precedent to warranty coverage.
2. If you have any questions about a Denver Instrument product, please contact the nearest Denver Instrument office as listed below.
3. If it becomes necessary to return your Denver Instrument product for service, you must obtain a "Return Authorization Number". Please pack the product securely in its original approved packing carton or an other suitable container. Include your Return Authorization Number on the shipping label. Shipping charges must be fully prepaid.

Return to authorized distributor or :

*North and South America:*

**Denver Instrument Company**  
6542 Fig Street  
Arvada, Colorado 80004  
1-800-321-1135  
Tel: 303-431-7255  
Fax: 303-423-4831

*U.K. and Ireland:*

**Denver Instrument Company**  
Denver House, Sovereign Way  
Trafalgar Business Park  
Downham Market  
Norfolk PE38 9SW England  
Tel: 44 136 63862 42  
Fax: 44 136 63862 04

*Europe, Asia and Australia:*

**Denver Instrument GmbH**  
Robert-Bosch-Briete 10  
37079 Gottingen Germany  
Tel: 49 551 20977 31  
Fax: 49 551 20977 39

# Maintenance and Troubleshooting

## Display Shows:




(Blank Screen)

Cause:	Remedy:
I/O key pressed and display is off.	Press the I/O key.
Power cord not connected.	Connect cord.
No power to outlet or improper voltage.	Check power supply and voltage switch.
Temporary fault	Disconnect and reconnect power cord. (Wait at least five seconds before reconnecting.)




(No stability indicator)

Weight exceeds balance capacity.	Reduce weight.
Pan obstructed.	Check pan placement.
Balance operating error.	Re-calibrate balance.
Load cell damaged.	The most common cause is from objects being dropped onto the weighing pan or from the balance itself being dropped. The balance must be returned for service.



(No stability indicator)


Air movement around balance.	Move balance to stable, draft-free location.
------------------------------	--




Usually detected when the calibration operation is being attempted. It is usually the result of the balance being exposed to vibrations, air drafts or unstable environments.	The balance must be placed in a stable environment before calibration occurs.
---	---



Pan was not in place when unit was turned on.	Turn unit off and place pan on and turn unit on.
---	--



This is an indication that an electronic malfunction has occurred.	The unit must be returned to the factory for evaluation.
--	--



This fault will appear on the display when the unit is powered on. This occurs when the electronics are no longer within the factory set parameters.	The most common cause is from objects being dropped onto the weighing pan or from the balance itself being dropped. The balance must be returned for service.
--	---

# Introduction

Thank you for choosing one of our precision instruments. Your balance is designed and manufactured to the most rigorous standards in order to give you years of service. First, check the contents of the shipping carton. You should find the following:

- Manual
- Balance
- AC Adapter
- Cover Pan

Next, follow the instructions for installing your balance.

Now you're ready to begin using your balance. To take advantage of its many features, carefully read your operating manual. It contains step-by-step procedures, examples, and other vital information.

Finally, remember to return your completed warranty card within ten days and to record all purchase information.



**Warning: Use of this product in a manner not specified by the manufacturer may impair any safety protection provided by the equipment!**

# Specifications

Model	XP-300	XP-600	XP-1500	XP-3000
Range	300g	600g	1500g	3000g
Readability	0.01g	0.02g	0.05g	0.1g
Linearity	±0.01g	±0.02g	±0.05g	±0.1g
Pan Dimensions	5" x 7" (127 x 178mm)	5" x 7" (127 x 178mm)	5" x 7" (127 x 178mm)	5" x 7" (127 x 178mm)
Permissible Calibration Weights	200g	500g	1000g	2000g
Temperature	15 - 40 °C			
Humidity	80% for temperature to 31°C, decreasing linearly to 50% relative humidity at 40°C			
Altitude	3000m			

**Common Specifications:** Unit dimensions (L x W x H) 8.2 x 8.3 x 2.7" (208 x 211 x 69mm); Net Weight: 3.75 lbs. (1.7 kg). Power: 15 VDC, 100mA, AC adapter included. Mains supply voltage fluctuations not to exceed ±10% of nominal supply voltage. Equipment id suitable for continuous operation with AC adapter/charger. Pollution degree:2; Installation category: II; Sound Pressure Level emitted by equipment does not exceed ambient noise. Storage/Transport Temperature: -20° – +40°C.

**NOTE: The battery is not a user serviceable part. If the battery does not recharge with use of AC adapter, contact an authorized service technician.**



**Caution: Use AC adapter/charger supplied with unit only! Consult Denver Instrument Company for replacement.**


## Preparation

- This product is intended for indoor use and portable applications.
- Select a suitable work area.
- Work area should be relatively free from drafts and vibrations.
- Work surface should be level and rigid.
- Do not locate near magnetic materials or equipment/instruments which use magnets in their design.
- Avoid areas which have variations in room temperatures or have excessive room temperatures. Room temperatures above 105°F/40°C or below 60°F/15°C could affect balance operation and accuracy.

## Installation

- Remove balance and accessories from the carton. Save packing material for transportation purposes.
- A rechargeable battery is already installed in your balance.
- Charge balance for 12 hours before using on battery power.
- When the battery icon flashes, you must recharge the unit.
- If using the AC adapter, insert power cord into the receptacle located on the rear of the unit. Firmly push in the plug.
- Allow the balance to warm up for 30 minutes prior to use.

Your balance features a numeric display that continuously shows your weighing results. The number of decimal places displayed depends on the XP model.



## Function Keys

- I/O** Turns balance ON or OFF, or Standby if AC adapter is connected.
- ↓** Scroll; for selecting various units of weight or options.
- \*** Print/Menu; press to print or hold down for more than 3 seconds to enter menu
- 0-** Zero/Tare; returns display to zero, exits from menu.

## Hardware Handshake

With a 4-wire interface, 1 or 2 more characters can be transmitted after CTS.



**Note:**

These connections must be made when interfacing the balance through the RS232C port!

### Data Interface Cable Pin Configuration For XP-Series Balances

DI XP Series Balance 9-Pin port		Standard RS-232 9-Pin Connector
RxD 2	←→	3 TxD
TxD 3	←→	2 RxD
DTR 4	←→	4 DTR
Signal Ground 5	←→	5 Signal Ground
CTS 8	←→	8 CTS

---

DI XP Series Balance 9-Pin Port		Standard RS-232 25-PIN Connector
RxD 2	←→	2 TxD
TxD 3	←→	3 RxD
DTR 4	←→	20 DTR
Signal Ground 5	←→	7 Signal Ground
CTS 8	←→	5 CTS

## RS232C Interface (Option)

The RS232C interface is an option added at the factory at the time of ordering your XP balance.

### Data Input Format

You can connect a computer or printer to your balance to send commands via the balance interface port to control balance functions and application.

A control command can have up to 4 characters. Each character must be transmitted according to the communication port setting for data transmission.

### Communications Port Settings

Baud	Data Bits	Parity	Stop Bit
9600	7bit	Space	1

Flow Control = None or Xon/Xoff

Line Delay 50ms

Character Delay 50 ms

### Format for Control Commands

Format: Esc	!	CR	LF
-------------	---	----	----

Esc: Escape

!: Command character

CR: Carriage Return (optional)

LF: Line Feed (optional)

Command character:	!	Meaning
	K	Very Stable Conditions
	L	Stable Conditions
	M	Unstable Conditions
	N	Very Unstable Conditions
	O	Disable Keypad
	P	Print Display
	R	Enable Keypad
	S	Reset Balance
	T	Tare and Zero (combined)
	U	Tare (tare only)
	V	Zero
	W	External Calibration/Adjustment


## Operation

Your new balance weighs in different units and can perform parts counting and percent of weight functions.

### A. Taring (Zeroing)

All models have taring capabilities up to their total weight capacity (varies with each model).

1. Place sample container on pan.
2. Press the **-0-** key to zero the balance.
3. Add your sample, when the balance is stable the display is showing the weight of the sample.



### B. Weigh Units

1. To toggle between pennyweight, ounces, grams, counting features (count or percentage depending on which one is activated) and optional unit (\*\*\*) , press the **↓** key.
2. To change the optional unit, press and hold down the **\*** key for 3 seconds. The display will read "Unit".
3. Press the **↓** key to cycle through the different weight modes.
4. Press the **-0-** key when the desired mode appears.

Lb	Pounds	GRAIN	Grains
Tr 02	Troy Ounces	CARAT	Carats
LAEL	Tael, Hong Kong		

